



AN ITALIAN COMPANY THAT RESEARCHES, DEVELOPS AND PRODUCES READY TO USE KITS  
USING CHROMATOGRAPHIC TECHNIQUES FOR CLINICAL LABORATORIES

## CONTROL LYOPHYL. IN URINE FOR 3-METHYLINDOLE - LEVELS 1 and 2 Code Z60019

### OBJECTIVE

These controls in plasma are used for the internal quality control and serve to monitor the accuracy and precision of analytical procedures devoted to the quantitative determination of analytes contained therein. These controls are lyophilized human matrix and are available in two different concentration ranges. Should be handled as if they were a real patient sample.

### RECOVERY




Remove the metal seal and rubber stopper from the vial. Add exactly 1 ml of HPLC-grade water through the tube. Replace the rubber stopper, allow to stand for at least 5 to 10 minutes. Before use, mix by inverting the vial to dissolve the material until a homogeneous solution.

### STORAGE AND STABILITY

The controls are stable for 12 months from the date of preparation if stored at -20 ° C. After reconstitution are stable for 1 day at -20 °C. Do not use after the expiry date.

### PRECAUTIONS



These controls in human matrix should be treated with care and treated as potentially infectious.

		
063	Oct 2023	Oct 2024

### PACKAGINGS AVAILABLE:

- Z60019 CONTROL IN URINE FOR 3-METHYLINDOLE (Skatole) - LEVELS 1 and 2      2 x 5 x 1 ml

CONCENTRATIONS  
CODE Z60019

LOT		
063	Oct 2023	Oct 2024

ANALYTE	MEASUREMENT UNIT	MEDIUM VALUE L1	RANGE L1	MEDIUM VALUE L2	RANGE L2
3-METHYLINDOLE	µg/l	15,3	10,7 – 19,9	75,4	52,8 – 98,1

This product fulfills all the requirements of Directive 98/79/EC of 27/10/1998 on in vitro diagnostic medical devices (IVD). The declaration of conformity is available upon request.



**FOR IN VITRO DIAGNOSTIC USE ONLY**

produced by



AN ITALIAN COMPANY THAT RESEARCHES, DEVELOPS AND PRODUCES READY TO USE KITS  
USING CHROMATOGRAPHIC TECHNIQUES FOR CLINICAL LABORATORIES

EUREKA srl – LAB DIVISION ,  
Via Enrico Fermi 25, 60033 Chiaravalle (AN) ITALY  
Tel. + 39 071 74 50 790  
Fax. + 39 071 74 96 579  
[info@eurekaone.com](mailto:info@eurekaone.com)  
[www.eurekakit.com](http://www.eurekakit.com)



COMPANY WITH  
MANAGEMENT SYSTEM  
CERTIFIED BY DNV  
= ISO 9001 =  
= ISO 13485 =

Release N° 002

September 2019